

(Försäkran om överensstämmelse)

Manufacturer

AKLA AB (Enhagsslingan 2) Box 534 182 15 Danderyd

Sweden

Sweden

Plum Quick Clean (955422A)

REF: 94407ST Wash swab

We hereby declare that the above mentioned devices comply with the Swedish legislation LVFS2003:11 transposing European Medical Devices Directive 93/42/EEC.

Product Class Is, Rule 1.

Intended use: Swab for cleansing of minor wounds.

EC certificate 41310145 valid to 2024-05-26 under the transition period to MDR2017/745

Täby 2021-03-01 AKLA AB

Eva Janmark Managing Director



(Försäkran om överensstämmelse)

Manufacturer AKLA AB

(Enhagsslingan 2)

Box 534

SE-182 15 Danderyd

Sweden

REF: 92111 Plum QuickStop Wound Dressing (955440)

Classification Class I Sterile, Rule 4
First Aid products for wounds.
Intended to be used as a mechanical barrier and compression

We hereby declare that the above mentioned devices comply with the Swedish legislation LVFS2003:11 transposing European Medical Devices Directive 93/42/EEC

EC certificate 41310145 valid to 2024-05-26 under the transition period to MDR2017/745

Täby 2021-02-24 AKLA AB

Eva Janmark

Managing director



(Försäkran om överensstämmelse)

Manufacturer AKLA AB

(Enhagsslingan 2)

Box 534

SE-182 15 Danderyd

Sweden

REF: 92109 Plum QuickStop Wound Dressing (955441)

Classification Class I Sterile, Rule 4
First Aid products for wounds.
Intended to be used as a mechanical barrier and compression

We hereby declare that the above mentioned devices comply with the Swedish legislation LVFS2003:11 transposing European Medical Devices Directive 93/42/EEC

EC certificate 41310145 valid to 2024-05-26 under the transition period to MDR2017/745

Täby 2021-02-24

AKLA AB

Eva Janmark Managing director



Manufacturer AKLA AB

(Enhagsslingan 2)

Box 534

SE-182 15 Danderyd

Plum QuickCool Burn Gel (955430)

REF: 94370 FOR BURNS Burn Gel 3,5 gram

Classification: Class IIa Rule 4.

FOR BURNS gel is a first aid product for burns. It is intended to be used with small burns. It is not being intended for burns and wounds requiring healing by secondary intent.

We hereby declare that the above mentioned devices comply with the Swedish legislation LVFS2003:11 transposing European Medical Devices Directive 93/42/EEC

EC certificate 41314724 valid to 2024-05-27 under the transition period to MDR2017/745

Täby 2021-02-24 **AKLA AB**

Eva Janmark Managing Director

Postadress Mailing adress AKI A AB Box 534

SE-182 15 DANDERYD

Fax. +46(0)583 71 15 05

Bank Bankgiro SEB 295-8049 Org.nr/VAT No

SE556027643701

Postgiro 1670-9



Manufacturer

AKLA AB

Enhagsslingan 2

SE-18740 Täby, Sweden

SRN (Single Registration Number): SE-MF-000004567

First Aid products for wounds. Classification Class I Sterile, Rule 4.

Intended use: As a mechanical barrier and compression.

REF: 92109 Pull1Aid Blood Stopper Mini (compress), First Aid bandage 7,5 x 12 cm
REF: 92111PM Pull1Aid Blood Stopper (compress), First Aid bandage 17 x 17 cm

We hereby declare that the above-mentioned devices comply with the European Medical Devices Directive 93/42/EEC

EC certificate 41310145 valid to 2024-05-26 under the transition period to MDR2017/745

Notified Body Intertek, 0413

Täby 2022-08-29 AKLA AB

Eva Janmark Managing director





州市富施达医疗器械有限公司 FIRSTAR HEALTHCARE Co., Ltd. (Guangzhou)



Manufacturer:

FIRSTAR HEALTHCARE COMPANY LIMITED (GUANGZHOU)

Address:

Rm. 901, Building No.2, Headquarters Center, Tian'an High-tech EcologicalPark, Panyu.

511400 Guangzhou, PEOPLE'S REPUBLIC OF CHINA

SRN:

Not available yet

European:

MedPath GmbH

Representative: Mies-van-der-Rohe-Strasse 8, 80807 Munich, Germany

SRN:

DE-AR-000000087

Trade name:

CPR Face Shield Product Name: CPR Face Shield

Product code: FS-100,FS-101,FS-102,FS-103,FS-105,FS-106

Applied Standard & Common Specification:

EN ISO 14971:2019

Basic UID:

Not available yet

Classification acc. to MDR Ax .VIII:

Class I, Rule 1

Conformity assessment procedure:

Annex II +Annex III of MDR

CE certificate No:

N.A

Name and ID of the Notified Body:

N.A

We, the manufacturer, herewith declare under our sole responsibility that the above-mentioned products meet the provisions of the Regulation (EU) 2017/745 on Medical Devices (MDR). All supporting documentations are retained under the premises of the manufacturer.

Signature

Name: Janice-Li

Position: MR

Place: Guangzhou

Date: 2021.4.2



Manufacturer AKLA AB

Enhagsslingan 2 SE-18740 Täby Sweden

REF. No Description Size Risk Class

REF 94373 Burn and Wound dressing 10x10cm Class Ilb Rule 4. Sterile

The dressings are first aid products for burns injuries and wounds. It is intended to be used with burns which have breached the dermis and can only heal by secondary intent.

We hereby declare that the above mentioned devices comply with the European Medical Devices Directive 93/42/EEC.

Notified Body

Intertek 0413

EC certificate 41314724-02 valid to 2024-05-27 under the transition period to MDR2017/745.

Täby 2022-08-29 AKLA AB

Eva Janmark

Managing Director



CE DECLARATION OF CONFORMITY

Plum Eyewash 0.9% Sodium Chloride

Eye irrigation fluid for first aid

We, the manufacturer, of the products covered by this declaration, hereby declare, that the bottles with sterile irrigation fluid listed below, meet the requirements in the Council Directive 93/42/EEC of 14 June 1993, the Directive 2007/47 /EC and its relevant transposition into national laws of the member states into which we place the device. Plum Safety ApS is responsible for the Declaration of Conformity concerning the medical devices listed below:

Art No	Art Name	Volume
5160	Plum Eyewash 0.9% Sodium Chloride QuickRinse	20ml
4691	Plum Eyewash 0.9% Sodium Chloride	200ml
4604	Plum Eyewash 0.9% Sodium Chloride	500ml
4707	Plum Eyewash 0.9% Sodium Chloride	1000 ml
4861	Plum Eyewash 0.9% Sodium Chloride DUO	500ml
4800	Plum Eyewash 0.9% Sodium Chloride DUO	1000 ml

With the intended purpose:

Plum Eyewash 0.9% Sodium Chloride is to be used as a first aid product for eye irrigation.

We guarantee and declare that:

- 1. The device is a Class Is device according to Annex IX, rule 5.
- 2. The irrigation fluid is STERILE
- 3. The planning and manufacture procedures are conducted in conformity with the Company Quality System, in compliance with the provisions of Annex VII combined with Annex V to the Directive.

The notified body DNV Product Assurance AS (ID No.2460) certifies Plum Safety ApS. The latest EC Certificate No. 10000457813-PA-NA-DNK 2021-05-05.

Assens, 2021-05-06

Djuych

Regulatory Affairs and Quality Assurance Manager

