

**Spec code:** LOF-050NA-N-3CZ (LOF-C0150)  
**Product name:** Sempercure® edition  
**Date of issue:** 23/10/15

**GENERAL INFORMATION**

**Type:** single-use disposable examination or protective glove  
**Marking:** all information on dispenser box  
**Shape:** ambidextrous – straight fingers  
**Material:** Natural Rubber Latex (NRL)  
**Colour:** natural white  
**Inner:** online chlorinated/powder free  
**Outer:** no treatment  
**Cuff/Surface:** rolled cuff/fingertip textured

<b>Thickness:</b> <small>(double measured)</small>	<b>Finger:</b> min. 0.22 mm	<b>Palm:</b> min. 0.20 mm	<b>Cuff:</b> min. 0.16 mm
---	-----------------------------	---------------------------	---------------------------

<b>Residual powder:</b> <small>[according to EN 455-3]</small>	≤ 2 mg / glove	<b>Force at break:</b> <small>[according to EN 455-2]</small>	med. ≥ 6.0 N (during shelf life)
---	----------------	--	-------------------------------------

<b>Length:</b> <small>[according to EN 455-2]</small>	median 240 mm	<b>Elongation:</b> <small>[according to ASTM D3578]</small>	min. 650 % (before aging) / min. 500% (before aging)
--	---------------	--	---


<b>Available Sizes:</b>	XS (5-6)	S (6-7)	M (7-8)	L (8-9)	XL (9-10)
-------------------------	----------	---------	---------	---------	-----------

**Storage:** 3 years storage life (date of expiration see on the box). Keep area cool, dry and dust free, avoid ventilation and storage close to photocopy equipment. Copper-ions discolour the glove. Protect gloves against ultraviolet light sources, as sunlight and oxidizing agents. Storage above 30°C will lead to accelerated aging and should be avoided.

**STANDARDS, GUIDELINES AND QUALITY CERTIFICATION**

**Quality certification:** ISO 9001, ISO 13485  
**Conformity to guidelines:** Personal Protective Equipment 89/686/EEC: CE cat. III  
 Medical Device Directive 93/42/EEC: CE Class I

**Conformity to standards:** EN 420 EN 374-2: specified EN 374-3: Class 4  
 EN 455 level 2 - AQL 1.5 Sodium Hydroxide 40%  
 ASTM D3578 level 3: AQL <0.65 (G1)  
 (except stress at 500% level 2: AQL <1.5 (G1)  
 elongation) level 1: AQL <4.0 (S4)




**Conformity to food regulation:** 1935/2004 

**WARNING AND REPORTING SYSTEM**

**Potentially allergenic ingredients:** Dithiocarbamate type, Zinc-mercaptobenzothiazol  
*(For further information, the actual version of the ingredients list is given on request.)*

**Medical Device Vigilance and Reporting System:** According to the official reporting criteria of MDD, incidents caused by examination gloves or surgical gloves must be reported immediately to our medical device reporting officer. Tel.: +43 2630 310 ext. -510; Fax: -549

**Caution!** This product contains accelerators and natural rubber latex which may cause allergic reactions, including anaphylactic responses!

Established (PM):	Reviewed (RA/R&D):	Approved (HoM/HoPM):
 L. Millonig	 Dr. R. Schaller	 Mag.(FH) K. Miller