### **MERCATOR** MEDICAL



# nitrylex<sup>®</sup> classic

The instruction below should be used in conjunction with detailed information on the packaging.

#### Short description of the product

Nitrile examination and protective gloves, powder-free, non-sterile for disposable use

#### Full description of the product

Raw material	: nitrile
External surface	: bisque with fingertip textured, polymerized
Internal surface	: polymerized + chlorinated
Cuff	: beaded
Colour	: blue/white/violet
Shape	: ambidextrous, fitting to the right and left hand
Size range	: XS (5-6), S (6-7), M (7-8), L (8-9), XL (9-10)
AQL	: 1.0
Quantity in packaging	: 50/100/200 pcs. by weight
Shelf life	: 3 years (from the date of manufacturing)

#### Storage instructions

It is recommended to store the gloves in dry place, in the temperature of 5-35°C and to protect them against direct sunlight and fluorescent light. Recommended relative humidity in the room where the gloves are stored is  $60 \pm 20\%$ .

Keep the gloves in a distance of not less than 1m from heating devices, sources of fire and ozone.

Do not keep in direct vicinity of solvents, oils, fuels and lubricants.

#### Food contact

Gloves are marked with food contact symbol  $\Im$  and comply with the requirements of Regulation (EU) No 10/2011, European Regulation (EC) No 1935/2004 and with Regulation (EC) No 2023/2006 on Good Manufacturing Practice. Gloves are suitable for handling any type of food and have been tested for Overall Migration Test acc. EN 1186:

Extraction conditions (tested for 2 h in 40°C)	Analysis results [mg/dm²]	Test Result (limit < 10 mg/dm²)		
3% acetic acid	1,1	Pass		
10% ethanol	<1	Pass		
Olive oil	<3	Pass		

MDD classification & compliance

Gloves are classified as class I Medical Device as per Annex IX of the Council Directive 93/42/EEC and comply to standards: EN 455-1:2000, EN 455-2:2015, EN 455-3:2015, EN 455-4:2009, EN ISO 15223-1:2016, EN 1041:2008+A1:2013.

#### PPE classification & compliance

Gloves are category III Personal Protective Equipment as per Annex I of the Regulation 2016/425 and comply to standards:

EN 420:2003+A1:2009, EN ISO 374-1:2016 (Type B), EN 374-2:2014, EN 16523-1:2015, EN 374-4:2013, EN ISO 374-5:2016.

Declaration of Conformity can be found under below web address: http://mercatormedical.eu/produkty/rekawice/diagnostyczne/nitrylex-classic

Notified Body 2777 responsible for EU Type Examination (Module B) and Module C2 On-going Conformity: Satra Technology Europe Ltd Bracetown Business Park, Clonee Dublin 15, Dublin, Ireland

CE2777

Intended use These are non-sterile examination and protective gloves for single use, intended for use in medical field to: protect patient and user from crosscontamination, conducting medical examinations, diagnostic and therapeutic procedures and for handling medical contaminated material. Gloves are classified as Medical Devices Class I and as a Personal Protective Equipment category III. Their design and labelling corresponds to the requirements of the European Medical Device Directive 93/42/EEC and the European Regulation 2016/425 on Personal Protective Equipment. Gloves should be used solely according to their intended application.

#### Precautions and indications for use

Dry hands before putting the gloves on. Before usage, inspect the gloves for any defect or imperfections. Use at least 1 pair of gloves for one patient and one procedure, these are disposable gloves. Do not let chemical substances get under the gloves through the cuff. If a chemical substance reaches the skin, wash it away immediately with plenty of water with soap. If the gloves get punctured, torn or broken during their use, take them off and put on the new ones. Avoid using gloves dirty in the inside as they may cause irritation leading to skin inflammation or more serious damages. The gloves should not be used in contact with open fire and to protect against any sharp tools. The gloves are not intended for welding, electric shock protection, ionizing radiation or from the effect of hot or cold objects.

This information does not reflect the actual duration of protection in the workplace and the differentiation between mixtures and pure chemicals.

The chemical penetration resistance has been assessed under laboratory conditions from samples taken from the palm only (except in case where glove is equal to or over 400 mm – where the cuff is tested also) and relates only to the chemical tested and to the tested specimen. It can be different if the chemical is used in a mixture.

It is recommended to check that the gloves are suitable for the intended use because the conditions at the workplace may differ from the type test depending on the temperature, abrasion and degradation.

When used, protective gloves may provide less resistance to the dangerous chemical due to changes in physical properties. Movements, snagging, rubbing, degradation caused by the chemical contact etc. may reduce the actual use time significantly. For corrosive chemicals, degradation can be the most important factor to consider in selection of chemical resistant gloves.

Gloves are suitable for special purposes as they are examination gloves where risk of injury to the wrist is considered to be minimal, gloves are shorter than EN 420 min. length requirement.

#### Components / hazardous components

Some gloves may contain components known to be a possible cause of allergy for person allergic to them, who may develop contact irritation and/or allergic reaction. In case of an allergic reaction, seek medical assistance immediately.

#### Disposal

Used gloves can be contaminated with contagious or other hazardous substances. They should be disposed of in accordance with local regulation. Gloves should be buried or burned under controlled conditions.

Manufacturer MERCATOR MEDICAL S.A

ul. H. Modrzejewskiej 30 31-327 Cracow, Poland www.mercatormedical.eu



## **MERCATOR** MEDICAL



<ul> <li>Level 1 &gt; 10 min • Level</li> </ul>	2 > 30 mir	• Level 3 > 60 min	<ul> <li>Level 4 &gt; 120 min • Level 5 &gt; 240 min • I</li> </ul>	Level 6 > 480 min	
Test results acc. to EN 16523-1:2015		EN 374-4:2013	Test results acc. to EN 16523-1:2015		EN 374-4:2013
Chemical	Level	Degradation [%]	Chemical	Level	Degradation (%)
*4% Chlorhexidine Digluconate	6	19.0	30% Hydrogen Peroxide (P)	2	22.8
40% Sodium Hydroxide (K)	6	-42.9	1.5% Methanol in water	6	21.9
10-13% Sodium Hypochlorite	6	14.7	25% Ammonium Hydroxide (O)	1	-52.0
50% Sulphuric Acid	6	-20.5	3% Povidione-iodine	6	33.7
10% Acetic Acid	4	66.7	10% Sodium Percarbonate	6	15.4
5% Ethidium Bromide	6	3.4	50% Glutaraldehyde	6	27.4
37% Formaldehyde (T)	3	5.0	0.1% Phenol	6	33.8

37% Formaldehyde (T)		3	5.0	0.1% Phenol		6	33.8		
*Permeation rate 7µg/c	m²/min, EN 374-4:2013 De	e change in puncture resistance of t	he gloves after ex	kposure t	o the challenge chemical.				
Test acc. To EN 374-2:2014 – Level 2 (ISO 2859)			Test acc. To EN ISO 374-5:2016						
Performance level AQL			*	Protection against bacteria & fungi Protection against viruses		Pass			
	Level 3 < 0.65 Level 2 <1.5		Protection	i agairist viruses	Pass				
	Level 1 < 4.0								
Symbols used on the packaging									
(2)	Do not re-use / gloves a intended for single use	ire	NON STERILE	Non-sterile gloves		Powde	red gloves		
	Do not use, if package i damaged	s	×	Keep away from solar and fluorescent light	Ň	Powde	r free gloves		
Ĵ	Keep away from moistu store in a dry place	ire,	5°C	Temperature limitation / gloves store in temperature 5- 35°C	POLYMER COATED		ce of polymer coating on er surface of the glove		
LATEX	Raw material – natural latex	rubber	<b>Q</b> ₃	Keep away from ozone	COSMETIC COATING		ce of cosmetic coating on er surface of glove		
REF	Catalogue number		LOT	Lot / batch number	SINGLET OXYGEN 102	Gloves oxygen	with incorporated singlet e layer.		
<b>EC REP</b>	EU Authorized Represe symbol should be accor by name and addre Authorized Represen	mpanied ss of	$\sum$	Expiry date	TEXTURED	Presen the glo	ce of external texture on ve		
	Marking of gloves prote against bacteria and fur	•		Gloves protecting against chemical dangers with digit literal odes	NITRYL	Gloves	made from nitrile		
VIRUS	Marking of gloves prote against viruses, bacteria fungi.	-	4	Antistatic gloves	VINYL	Gloves	made from vinyl		
ISO 374-1:2016/Type A	Marking o type A chemical gloves. Six tested chemical identified by their code let pictogram.	s shall be		Date of manufacture	NEOPREN	Gloves	made from neoprene		
150 3/4-1.2016/ Туре В ХУZ	Marking o type B chemical gloves. Three tested chemi be identified by their code under pictogram.	cals shall		Manufacturer, symbol should be accompanied by name and address of Manufacturer	POLIIZOPREN	Gloves	made from polyisoprene		
ISO 374-1:2016/Type C	Marking o type C chemical gloves. One tested chemica be identified by their code under pictogram.	als shall letter	(11)	Consult instructions for use	M 50 by weight	50 glov	es by weight		
ŀ	Protective glove against me risk (if applicable accompar digit code of relevant perfo levels)	nied by 4	PAP	Package made from paper, qualify for recycling	by weight	100 glc	ves by weight		

Food contact symbol (article is suitable for food contact, for details check the instruction for use)

Indicates compliance with the requirements of Russian market

municipal waste Indicates compliance with the

Package is treated as

requirements of Ukrainian market 200 gloves by weight Additonal information on inner

side of package

User instructions Rev. 1.7, October 2019

<u>[]i</u>]

200

by weight

1 D

www.mercatormedical.eu





Hold the glove with a finger of the other hand at the cuff.

Hold the other glove by its outer edge.

### ■ HOW TO PUT THE GLOVES ON? ■

**MERCATOR** MEDICAL







Done



