

## EU Declaration of Conformity

**Manufacturer:**

Xiantao Yi-Ya Protective Products Co.,Ltd.  
Mian dong Road, Hefeng Village, Xiliuhe Town,  
Xiantao City,HubeiProvince,China  
Tel +86 18696471067  
Daisy@xtvya.com

**European**

Luxus Lebenswelt GmbH

**Representative:**

Kochstr. 1, 47877, Willich, Germany

**DIMDI Code**

DE/0000047791

**Tax Number**

DE305829099

**Contact Person**

Lin Sun

**Tel/Fax**

0049-1715605732

**E-mail**

Info.m@luxuslw.de

**Product Name**

Disposable face mask/Disposable protective mask

Disposable Medical face mask

**Test Article**

YY-1, YY-2, YY-3, YY-4, YY-5

Classification(MDD, Annex VIII) : Class 1,Rule 1

We herein declare that the above mentioned products meet the transposition into national law, the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer:it bears the mark



General applicable regulations, directives:

Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices, amending Directive 2001/83/EC, Regulation (EC) No 178/2002 and Regulation (EC) No 1223/2009 and repealing Council Directives 90/385/EEC and 93/42/EEC.

Applied standards, common specification, guidance:


EN 14683:2019+AC:2019, EN ISO 15223-1:2016, EN 1041:2008, EN ISO 14971:2012, EN62366-1:2015+AC:2015, ISO 10993- 1:2018, ISO 10993-5:2009, ISO 10993-10:2010, ASTM D4169-2016, MDCG 2019-15.

13<sup>th</sup> Dec,2019

Zhao Junpeng

General Manager

Xiantao Yi-Ya Protective Products Co.,Ltd.



## Synthetic Blood Penetration Resistance Final Report

Test Article: Disposable medical mask - 1  
Disposable medical mask - 2  
Disposable medical mask - 3  
Disposable medical mask - 4  
Disposable medical mask - 5

Study Number: 1289150-S01.1 Amended

Study Received Date: 16 Apr 2020

Study Completion Date: 27 Apr 2020

Testing Facility: Nelson Laboratories, LLC  
6280 S. Redwood Rd.

Salt Lake City, UT 84123 U.S.A.

Test Procedure(s): Standard Test Protocol (STP) Number: STP0012 Rev 09

Deviation(s): None

**Summary:** This procedure was performed to evaluate surgical facemasks and other types of protective clothing materials designed to protect against fluid penetration. The purpose of this procedure is to simulate an arterial spray and evaluate the effectiveness of the test article in protecting the user from possible exposure to blood and other body fluids. The distance from the target area surface to the tip of the cannula is 30.5 cm. A test volume of 2 mL of synthetic blood was employed using the targeting plate method.

This test method was designed to comply with ASTM F1862 and ISO 22609 (as referenced in EN 14683:2019 and AS4381:2015) with the following exception: ISO 22609 requires testing to be performed in an environment with a temperature of  $21 \pm 5^\circ\text{C}$  and a relative humidity of  $85 \pm 10\%$ . Instead, testing was performed at ambient conditions within one minute of removal from the environmental chamber held at those parameters.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Number of Test Articles Tested: 32 per test pressure

Number of Test Articles Passed: 32

Test Side: Outside

Pre-Conditioning: Minimum of 4 hours at  $21 \pm 5^\circ\text{C}$  and  $85 \pm 5\%$  relative humidity (RH)

Test Conditions:  $22.2^\circ\text{C}$  and 21% RH



*Shuang Shuang for*  
Study Director

James W. Luskin

Amended Report Date

*07 May 2020*



1289150-S01

**Results:** Per ASTM F1862 and ISO 22609, an acceptable quality limit of 4.0% is met for a normal single sampling plan when  $\geq 29$  of 32 test articles show passing results.

Test Pressure: 120 mmHg (16.0 kPa)

Test Article Number

1-32

Synthetic Blood Penetration

None Seen

**Amendment Justification:** At the request of the sponsor, the company name and address were updated. Additionally, the initial report was separated into individual reports by test pressure.

## Microbial Cleanliness (Bioburden) of Medical Masks Final Report

---

Test Article: Disposable medical mask -1  
Disposable medical mask - 2  
Disposable medical mask - 3  
Disposable medical mask - 4  
Disposable medical mask - 5

Study Number: 1289157-S01

Study Received Date: 16 Apr 2020

Testing Facility: Nelson Laboratories, LLC  
6280 S. Redwood Rd.

Salt Lake City, UT 84123 U.S.A.

Test Procedure(s): Standard Test Protocol (STP) Number: STP0036 Rev 15

Customer Specification Sheet (CSS) Number: 202002111 Rev 01

Deviation(s): None

**Summary:** The testing was conducted in accordance with EN 14683:2019, with the exception of approximate volumes of eluent used when performing the extraction procedure and a temperature range of 30-35°C used for aerobic incubation.

When bioburden results are calculated using a software program, manual calculations may differ slightly due to rounding. The counts determined on products are colony forming units and may not always reflect individual microorganisms. The sponsor performs any statistical analysis and determines the acceptable limits. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.



---

Carl Danielson electronically approved for  
Study Director

Robert Putnam

29 Apr 2020 17:53 (+00:00)

Study Completion Date and Time



**Results:**

Unit Number	Weight (g)	Aerobic	Fungal	Total Bioburden (CFU/mask)	Total Bioburden (CFU/g)
1	3.2	<3	<3	<6.1	<1.9
2	3.1	<3	<3	<5.8	<1.9
3	3.1	<3	<3	<6.1	<2.0
4	3.1	<3	<3	<5.8	<1.9
5	3.1	<3	<3	<6.0	<1.9
Recovery Efficiency	UTD <sup>a</sup>				

< = No Organisms Detected

UTD = Unable to Determine

Note: The results are reported as colony forming units per test article.

<sup>a</sup> UTD due to zero count on the first rinse. An alternative method or inoculated product recovery efficiency is recommended.

Method Suitability:

Organism	Percentage
<i>Bacillus atrophaeus</i>	88%

**Test Method Acceptance Criteria:** If applicable, anaerobic controls are acceptable for the bioburden test results. The number of masks to be tested shall be a minimum of 5 or more to meet an acceptable quality level of 4%. The bioburden of the medical mask shall be < 30 CFU/g tested.

**Procedure:**

- Positive Controls/Monitors: *Bacillus atrophaeus*
- Extract Fluid: Peptone Tween<sup>®</sup>
- Extract Fluid Volume: ~300 mL
- Extract Method: Orbital Shaking for 15 minutes at 250 rpm
- Plating Method: Membrane Filtration
- Agar Medium: Potato Dextrose Agar  
Tryptic Soy Agar
- Recovery Efficiency: Exhaustive Rinse Method
- Aerobic Bacteria: Plates were incubated 3 - 7 days at 30-35°C, then enumerated.
- Fungal: Plates were incubated 5 - 7 days at 20-25°C, then enumerated.

## Bacterial Filtration Efficiency (BFE) and Differential Pressure (Delta P) Final Report

Test Article: Disposable medical mask -1  
Disposable medical mask - 2  
Disposable medical mask - 3  
Disposable medical mask - 4  
Disposable medical mask - 5  
Study Number: 1289151-S02  
Study Received Date: 16 Apr 2020  
Testing Facility: Nelson Laboratories, LLC  
6280 S. Redwood Rd.  
Salt Lake City, UT 84123 U.S.A.  
Test Procedure(s): Standard Test Protocol (STP) Number: STP0004 Rev 18  
Deviation(s): None

**Summary:** The BFE test is performed to determine the filtration efficiency of test articles by comparing the bacterial control counts upstream of the test article to the bacterial counts downstream. A suspension of *Staphylococcus aureus* was aerosolized using a nebulizer and delivered to the test article at a constant flow rate and fixed air pressure. The challenge delivery was maintained at  $1.7 - 3.0 \times 10^3$  colony forming units (CFU) with a mean particle size (MPS) of  $3.0 \pm 0.3 \mu\text{m}$ . The aerosols were drawn through a six-stage, viable particle, Andersen sampler for collection. This test method complies with ASTM F2101-19 and EN 14683:2019, Annex B.

The Delta P test is performed to determine the breathability of test articles by measuring the differential air pressure on either side of the test article using a manometer, at a constant flow rate. The Delta P test complies with EN 14683:2019, Annex C and ASTM F2100-19.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Side: Inside  
BFE Test Area:  $\sim 40 \text{ cm}^2$   
BFE Flow Rate: 28.3 Liters per minute (L/min)  
Delta P Flow Rate: 8 L/min  
Conditioning Parameters:  $85 \pm 5\%$  relative humidity (RH) and  $21 \pm 5^\circ\text{C}$  for a minimum of 4 hours  
Test Article Dimensions:  $\sim 177 \text{ mm} \times \sim 94 \text{ mm}$   
Positive Control Average:  $1.7 \times 10^3$  CFU  
Negative Monitor Count:  $< 1$  CFU  
MPS:  $2.9 \mu\text{m}$



Trang Truong electronically approved for  
Study Director

James Luskin

19 May 2020 00:42 (+00:00)  
Study Completion Date and Time

**Results:**

Test Article Number	Percent BFE (%)
1	99.8
2	99.6
3	99.9
4	>99.9
5	99.8

Test Article Number	Delta P (mm H <sub>2</sub> O/cm <sup>2</sup> )	Delta P (Pa/cm <sup>2</sup> )
1	4.4	43.6
2	4.2	41.5
3	4.3	42.1
4	4.7	46.4
5	4.2	41.4


The filtration efficiency percentages were calculated using the following equation:

$$\% BFE = \frac{C - T}{C} \times 100$$

C = Positive control average

T = Plate count total recovered downstream of the test article

Note: The plate count total is available upon request



中国认可  
国际互认  
检测  
TESTING  
CNAS L0599

Test Report                      **SL52025245933401TX**                      Date: May 15, 2020                      Page 1 of 3  
 XIANTAO JIAJING PROTECTIVE PRODUCTS CO.,LTD  
 7-17-301 XIANDAI ZHONGJIA SCIENCE AND TECHNOLOGY CITY, SHAZUI STREET, XIANTAO CITY

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description            : (A) Disposable face mask

SGS Internal Ref No.         : O-CQAFL202000392405

Sample Color                 : (A) Blue

Test Performed               : Selected test(s) as requested by applicant

Sample Receiving Date     : Apr 26, 2020

Testing Period                : Apr 26, 2020 - May 15, 2020

Test Result(s)                : Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of  
 SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/sgs/terms-and-conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/sgs/terms-and-conditions/terms-and-conditions-document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in the test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephones: (86-755)93071443, or e-mail: [CS@cschehold.com](mailto:CS@cschehold.com)

SGS-CSTC 标准技术服务有限公司 (Shanghai) Co., Ltd.      3 Building No.888 Yahan Road, Xuhai District, Shanghai, China      200233      1 (86-21) 61402866      1 (86-21) 64958783      [www.sgs.com](http://www.sgs.com)  
 中国·上海·徐汇区宜山路888号3号楼      邮编: 200233      1 (86-21) 61402866      1 (86-21) 64958783      [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

Member of the SGS Group (SGS SA)





Test Report

SL52025245933401TX

Date: May 15, 2020

Page 2 of 3

Test Result

**Medical Face Masks-Requirements and Test Methods**  
(EN 14683:2019)

**Clause 5.2.2 Bacterial filtration efficiency (BFE)\***  
(EN 14683 :2019 Annex B)

(BFE), %	1#	2#	3#	4#	5#
	99.6	99.6	99.6	99.8	99.7

Remark: Performance Requirement: Type I  $\geq 95\%$ , Type II  $\geq 96\%$ , Type IIR  $\geq 96\%$

\* This test standard is not within the accredited scope in SGS Shanghai testing centre, it is carried out by external laboratory accredited by CMA (China Metrology Accreditation).

**Clause 5.2.3 Breathability (Differential Pressure)**  
(EN 14683 :2019 Annex C, Flow rate 8 l/min)

Differential pressure $\Delta P$ (Pa/cm <sup>2</sup> )	1#	2#	3#	4#	5#
	34	34	37	36	36

Remark: Performance Requirement: Type I < 40 Pa/cm<sup>2</sup>, Type II < 40 Pa/cm<sup>2</sup>, Type IIR < 60 Pa/cm<sup>2</sup>

**Clause 5.2.4 Splash Resistance**  
(ISO 22609 :2004, Pressure 16.0 kPa)

Penetration on inside surface							
1#	2#	3#	4#	5#	6#	7#	8#
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
9#	10#	11#	12#	13#	14#	15#	16#
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
17#	18#	19#	20#	21#	22#	23#	24#
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
25#	26#	27#	28#	29#	30#	31#	32#
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
Number of Pass:	32						
Overall result:	Acceptable						

Remark:

- 1) Performance Requirement Type I: N/A, Type II: N/A, Type IIR:  $\geq 16.0$ kPa
- 2) Distance of the medical face mask target area surface to the tip of cannula is  $300 \pm 10$ mm.
- 3) Condition and Test temperature  $(21 \pm 5)^\circ$  C, relative humidity  $(85 \pm 10)\%$
- 4) An acceptable quality limit of 4.0% is met for a single sampling plan when 29 or more of the 32 tested specimens show pass results



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.cn/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/terms-and-conditions/terms-a-document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the samples tested and such samples are retained for 30 days only.

Attention: To check the authenticity of testing/inpection report & certificate, please contact us at telephone: (86-758) 3307 1443, or email: [CN\\_Service@sgs.com](mailto:CN_Service@sgs.com)  
 3<sup>rd</sup> Building, No. 888, Yahan Road, Xuhui District, Shanghai, China 200233 1 (86-21) 61402886 1 (86-21) 64958763 www.sgs.com.cn  
 中国·上海·徐汇区宜山路888号3号楼 邮编: 200233 1 (86-21) 61402886 1 (86-21) 64958763 \* [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

Member of the SGS Group (SGS SA)

**Clause 5.2.5 Microbial Cleanliness**  
(EN 14683: 2019 Annex D)

Sample A	1#	2#	3#	4#	5#
CFU/g	1	<1	<1	<1	<1

Remark: Performance Requirement: Type I≤30 CFU/g, Type II≤30 CFU/g, Type IIR≤30 CFU/g

Sample Photo



The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

\*\*\*End of Report\*\*\*



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic formal documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction clauses defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the reagents shown in this test report refer only to the sample(s) tested and such reagents are retained for 30 days only. Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: [CN\\_Quechek@sgs.com](mailto:CN_Quechek@sgs.com)

5<sup>th</sup> Building, No. 888, Yahan Road, Xuhai District, Shanghai, China 200233      1 (86-21) 61402556      1 (86-21) 61455753      [www.sgs.com.cn](http://www.sgs.com.cn)  
中国·上海·徐汇区宜山路888号3号楼      邮编: 200233      1 (86-21) 61402556      1 (86-21) 61455753      [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

# TEST REPORT

(Electronic version)



No:200076624

VERIFICATION WEBSITE: [www.gtgc.net.cn](http://www.gtgc.net.cn)

VERIFICATION CODE: MFSD-8322-54



ISSUE DATE:2020-04-20

APPLICANT: XIANTAO YI-YA PROTECTIVE PRODUCTS CO.,LTD  
ADDRESS: MIANDONG ROAD HEFENG VILLAGE XILIUHE TOWN XIANTAO CITY, HUBEI PROVINCE, CHINA

## INFORMATION CONFIRMED BY APPLICANT:

ORDINARY PROTECTIVE MASK  
QUANTITY: THIRTY-FIVE PIECES  
COLOUR: BLUE

DATE RECEIVED/DATE TEST STARTED: 2020-04-14

## CONCLUSION:

BACTERIAL FILTRATION EFFICIENCY	M
PARTICLE FILTRATION EFFICIENCY	M
AIRFLOW RESISTANCE	M

NOTE: "M" -MEET THE STANDARD'S REQUIREMENT "F" -FAIL TO MEET THE STANDARD'S REQUIREMENT  
"---" -NO COMMENT

## REMARK:

THE DECISION INDICATORS ARE DERIVED FROM THE GROUP STANDARD REQUIRED BY CLIENT (T/CTCA 7-2019). OUR INSPECTION CAPACITY AUTHORIZED BY CMA COVERS THE INSPECTION ITEMS T/CTCA 7-2019.  
THIS REPORT IS THE ENGLISH TRANSLATION VERSION OF THE REPORT 200076623.  
ALL THE TESTED ITEMS ARE TESTED UNDER THE STANDARD CONDITION (EXCEPT FOR INDICATION).  
COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.  
THE EXPERIMENT WAS CARRIED OUT AT No.1, ZHUJIANG ROAD, PANYU DISTRICT, GUANGZHOU, GUANGDONG, P. R. CHINA.

APPROVED BY:  
Yuan Liu ENGINEER

刘圆




1 OF 3

TEST REPORT

(Electronic version)

No:200076624





## TEST REPORT

(Electronic version)

No:200076624

### BACTERIAL FILTRATION EFFICIENCY(%)

(YY 0469-2011 ANNEX B, TEST BACTERIA: STAPHYLOCOCCUS AUREUS ATCC 6538, TEST AREA:  
40cm<sup>2</sup>, FLOW RATE: 28.3L/min, MEAN PARTICLE SIZE: 3.0 μm, RESULT OF THE POSITIVE  
CONTROL: 1.9×10<sup>8</sup> CFU, RESULT OF THE NEGATIVE CONTROL: <1CFU)

		REQUIREMENT
BFE <sub>1</sub>	99.6	≥95
BFE <sub>2</sub>	99.4	(T/CTCA 7-2019)
BFE <sub>3</sub>	99.2	

### PARTICLE FILTRATION EFFICIENCY(%)

(YY 0469-2011 5.6.2, AIR FLOW: 30L/min, AEROSOL: NaCl, AEROSOL CONCENTRATION:  
15mg/m<sup>3</sup>, TEMP: 23.2℃, RH: 36.3%)

		REQUIREMENT
MINIMUM	96.02	≥80
		(T/CTCA 7-2019)

### AIRFLOW RESISTANCE(Pa)

(YY 0469-2011 5.6.2, AIR FLOW: 30L/min, AEROSOL: NaCl, AEROSOL CONCENTRATION:  
15mg/m<sup>3</sup>, TEMP: 23.2℃, RH: 36.3%)

		REQUIREMENT
MAXIMUM	40.0	≤80
		(T/CTCA 7-2019)

——End of Report——



PAGE 3 OF 3



# Test Report

(Electronic version)

Verification Website: [www.gttc.net.cn](http://www.gttc.net.cn)

Verification Code: FVMR-0184-54

No: 20R000519

Issue Date: 2020-04-24

Applicant: XIANTAO YI-YA PROTECTIVE PRODUCTS CO.,LTD  
 Address: MIANDONG ROAD HEFENG VILLAGE XILIUHE TOWN XIANTAO CITY,HUBEI PROVINCE,CHINA

Information confirmed by applicant:

Disposable medical mask(non-sterile)

Quantity: sixty pieces

Lot number: 042020

Model: YY-101 YY-102 YY-103(submission no.: YY-101)

Size: 175mm×95mm

Classification: Type II R

Standard Adopted:

EN 14683:2019+AC:2019 <Medical face masks-Requirements and test method>

Date Received/Date Test Started: 2020-04-09

Conclusion:

Bacterial filtration efficiency (BFE)	M
Microbial cleanliness	M
Differential pressure	M
Splash resistance pressure	M
Materials and construction	M
Design	M
General	M

Note: "M"-Meet the standard's requirement "F"-Fail to meet the standard's requirement "--"-No comment

Remark:

All the tested items are tested under the standard condition (except for indication).

Copies of the report are valid only re-stamped.

The experiment was carried out at No 1, Zhujiang Road, Panyu District, Guangzhou, Guangdong, P.R.China.

Approved By:

*ZiShan Guo*

ZiShan Guo Senior Engineer



# Test Report

(Electronic version)

No: 20R000519



# Test Report

(Electronic version)

No: 20R000519

**Bacterial filtration efficiency (BFE)**

Test method: EN 14683: 2019+AC: 2019 Annex B

**Test principle:**

A specimen of the mask material is clamped between a six-stage cascade impactor and an aerosol chamber. An aerosol of *Staphylococcus aureus* is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency (BFE) of the mask is given by the number of colony forming units passing through the medical face mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

**Test equipment:**

- Incubator
- Electronic balance
- Autoclave
- Experimental system for bacterial filtration efficiency (BFE) of mask

**The environmental conditions of the laboratory and test condition:**

- Total bacteria: 0 CFU/plate
- Total fungi: 0 CFU/plate
- Blank experiment: Aseptic growth
- Test environment temperature: 24.5°C, Relative humidity: 56.0%
- Culture medium: TSA agar medium
- Culture temperature: 37°C, Culture time: 48h
- Test bacteria : *staphylococcus aureus* ATCC 6538
- Concentration of bacterium:  $5.0 \times 10^5$  CFU /ml
- Positive control average (C):  $1.9 \times 10^5$  CFU
- Negative monitor count: <1 CFU
- Test area: 40 cm<sup>2</sup>
- Dimensions of the test specimens: 15cm×15cm
- Flow rate: 28.3 l/min
- Pretreatment: Condition each specimen for 4 h by exposure to a temperature of (21±5)°C and a relative humidity of (85±5)%
- Mean particle size: 3.0 μm
- The medical face mask in contact with the bacterial challenge: inside



# Test Report

(Electronic version)

No: 20R000519

**Results:**

Sample	T	BFE (%)	Requirement (%)	Classification	Conclusion
1	9	99.53	≥98 EN 14683:2019+AC:2019	Type II R	Pass
2	8	99.58			
3	6	99.68			
4	11	99.42			
5	14	99.26			

**Remarks:**

For each test specimen calculate the bacterial filtration efficiency B, as a percentage, using the following formula:

$$B = (C - T) / C \times 100$$

where

B is bacterial filtration efficiency (BFE), %;

C is positive control average;

T is the total plate count for the test specimen.



# Test Report

(Electronic version)

No: 20R000519

## Microbial cleanliness

Test method: EN ISO 11737-1:2018, Membrane filtration

### Test principle:

Take the required samples from the original packaging. Weigh a certain amount of sample and placed in a sterile 500 ml bottle containing 300 ml of extraction liquid (1 g/l Peptone, 5 g/l NaCl and 2 g/l Tween 20). The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm. After this extraction step, 100 ml of the extraction liquid is filtered through a 0.45 µm filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 ml aliquot of the same extraction liquid is filtered in the same way and the filter plated on Sabouraud Dextrose agar (SDA) for fungi enumeration. The plates are incubated for 3 days at 30°C and 7 days at (20 to 25)°C for TSA and SDA plates respectively. The total bioburden is expressed by addition of the TSA and SDA counts.

### Test equipment:

Constant temperature incubator  
Electronic balance  
Pressure steam sterilizer  
Biosafety cabinet

### The environmental conditions of the laboratory and test condition:

Test environment temperature: 24.5°C, Relative humidity: 56.0%

Test environment monitoring: total bacteria: 0 CFU/plate, total fungi: 0 CFU/plate, blank experiment: aseptic growth





# Test Report

(Electronic version)

No: 20R000519

**Results:**

Microbial	Measured value (CFU/g)	Microbial cleanliness (CFU/g)	Requirement (CFU/g)	Classification	Conclusion
Bacteria	0	0	≤30 EN 14683:2019+AC:2019	Type II R	Pass
Fungi	0				



# Test Report

(Electronic version)

No: 20R000519

**Differential pressure**

Test method: EN 14683:2019+AC:2019 Annex C

**Test principle:**

This procedure was performed to evaluate the differential pressure of the medical face mask material by measuring the air exchange pressure through a measured surface area at a constant air flow rate.

**Test equipment:**

GTTC-YLC-1 Apparatus for measuring differential pressure

**The environmental conditions of the laboratory and test condition:**

Air flow: 8 l/min

Test area: 4.9cm<sup>2</sup>

Pretreatment: Condition each specimen for a minimum of 4 h by exposure to a temperature of (21±5) °C and a relative humidity of (85±5)%

General location of the areas of the mask the differential measurements: specimen center



# Test Report

(Electronic version)

No: 20R000519

**Results:**

Sample	Measured value (Pa)	Differential pressure (Pa/cm <sup>2</sup> )	Requirement (Pa/cm <sup>2</sup> )	Classification	Conclusion
1	167	33.7	< 60 EN 14683:2019+AC:2019	Type II R	Pass
2	163				
3	172				
4	150				
5	174				
Average	165				



# Test Report

(Electronic version)

No: 20R000519

## Splash resistance pressure

Test method: ISO 22609:2004

### Test principle:

A specimen medical face mask is supported on an apparatus. A volume of synthetic blood is sprayed horizontally at the specimen mask to simulate the scenario of a mask being splashed by a punctured blood vessel. The volume of fluid, distance to impact, orifice size and fluid velocity are defined in this method and intended to be consistent with this health care scenario. Any evidence of synthetic blood penetration on the side of the medical face mask contacting the wearer's face constitutes failure. Results are reported as "pass/fail". Specimen medical face masks are evaluated at a total of three different velocities corresponding to human blood pressures of 10.6 kPa, 16.0 kPa, and 21.3 kPa. Test results are reported at each velocity and the medical face mask is rated at the highest corresponding blood pressure for which medical face mask specimens demonstrate an acceptable quality limit of 4.0.

### Test equipment:

Test apparatus for synthetic blood penetration LFY-227

Air compressor

Graduated cylinder

Electronic balance

Targeting plate

### The environmental conditions of the laboratory and test condition:

Condition each specimen for a minimum of 4 h by exposure to a temperature of  $(21\pm 5)^{\circ}\text{C}$  and a relative humidity of  $(85\pm 5)\%$

Surface tension of synthetic blood: 0.042 N/m

Pressure: 16.0 kPa

Velocity: 550 cm/s



# Test Report

(Electronic version)

No: 20R000519

**Results:**

Sample	Measured value	Requirement (kPa)	Classification	Conclusion
	Pressure			
	16.0 kPa			
1	pass	≥16.0 EN 14683:2019+AC:2019	Type II R	Pass
2	pass			
3	pass			
4	pass			
5	pass			
6	pass			
7	pass			
8	pass			
9	pass			
10	pass			
11	pass			
12	pass			
13	pass			
14	pass			
15	pass			
16	pass			
17	pass			
18	pass			
19	pass			
20	pass			
21	pass			
22	pass			
23	pass			
24	pass			
25	pass			
26	pass			
27	pass			
28	pass			
29	pass			
30	pass			
31	pass			
32	pass			
<b>Final result</b>	pass			

**Remarks:**

An acceptable quality limit of 4.0 % is met for a single sampling plan when 29 or more of the show "pass" results.





# Test Report

(Electronic version)

No: 20R000519

**Materials and construction**

**Test Method:** EN 14683:2019+AC:2019 5.1.1

**Results:**

Requirement	Conclusion
The medical face mask is a medical device, generally composed of a filter layer that is placed, bonded or moulded between layers of fabric.	Pass
The medical face mask shall not disintegrate, split or tear during intended use.	Pass
In the selection of the filter and layer materials, attention shall be paid to cleanliness.	Pass



# Test Report

(Electronic version)

No: 20R000519

**Design**

Test Method: EN 14683:2019+AC:2019 5.1.2

**Results:**

Requirement	Conclusion
The medical face mask shall have a means by which it can be fitted closely over the nose, mouth and chin of the wearer and which ensures that the mask fits closely at the sides.	Pass
Medical face masks may have different shapes and constructions as well as additional features such as a face shield (to protect the wearer against splashes and droplets) with or without anti-fog function, or a nose bridge (to enhance fit by conforming to the nose contours).	Pass



# Test Report

(Electronic version)

No: 20R000519

**General**

Test Method: EN 14683:2019+AC:2019 5.2.1

**Results:**

Requirement	Conclusion
All tests shall be carried out on finished products or samples cut from finished products.	Pass



——End of Report——

