

**EU DECLARATION OF CONFORMITY**

Manufacturer: **MERCATOR MEDICAL S.A.**  
 UL. H.MODRZEJEWSKIEJ 30  
 31-327 KRAKÓW, POLSKA

Declares under its sole responsibility that non-sterile examination and protective gloves:

| Brand             | Type   | Sizes                | Reference Numbers   |
|-------------------|--|----------------------|---|
| nitrylex® classic | nitrile, powder-free, blue, for single use   | XS (5-6) - XL (9-10) | a'100: RD30019001-05<br>a'200: RD30096001-05                        |
|                   | nitrile, powder-free, white, for single use  | XS (5-6) - XL (9-10) | a'50: RD30174001-05<br>a'100: RD30143001-05<br>a'200: RD30097001-05 |
|                   | nitrile, powder-free, violet, for single use | XS (5-6) - XL (9-10) | a'100: RD30169001-05<br>a'200: RD30168001-05                        |

classified as medical device class I according to Annex IX of the Council Directive 93/42/EEC meet the essential requirements of Annex I of the Council Directive 93/42/EEC amended by the Directive 2007/47/EC and comply with the European harmonized standards: EN 455, EN ISO 15223-1, EN 1041. Conformity assessment procedure performed according Annex I and Annex VII of the Council Directive 93/42/EEC amended by the Directive 2007/47/EC.

The products described above are also classified as Personal Protective Equipment Category III and comply with Regulation (EU) 2016/425 of the European Parliament and the Council of 9 March 2016 on Personal Protective Equipment and resolution of the Council Directive 89/686/EEC and European standards: EN 420:2003+A1:2009, EN ISO 374-1:2016, EN 374-2:2014, EN 16523-1:2015, EN 374-4:2013, EN ISO 374-5:2016.

The products described above are identical to the Personal Protective Equipment, which is the subject to the EU Type Examination (Module B) under certificate No. 2777/10015-03/E17-01 issued by notified body:

**Satra Technology Europe Limited (2777)**  
 Bracetown Business Park, Clonee, Dublin 15, Dublin, Ireland

and are subject to the conformity to type procedure based on the internal production control plus supervised product checks at random intervals (Module C2) under surveillance of the notified body:

**Satra Technology Europe Limited (2777)**  
 Bracetown Business Park, Clonee, Dublin 15, Dublin, Ireland

Date and place of issue:  
 30.09.2019, Kraków

Signed on the behalf of the Manufacturer:



Wojciech Hercka  
 Product Documentation Manager

**MERCATOR MEDICAL S.A.**  
 ul. Heleny Modrzejewskiej 30, 31-327 Kraków  
 tel. 12 66 55 400, fax 12 66 55 415  
 Rejestracja: Sąd Rejonowy dla Krakowa - Śródmieścia w Krakowie,  
 XI Wydział Gospodarczy KRS, KRS: 0000036244  
 Kapitał zakładowy (w całości wpłacony): 10.589.100 PLN  
 NIP: 677-10-36-424, REGON: 350967107  
 Numer BDO: 000056063  
 -6-