

## PROSENSO PEARL (EWLTX58M)

### ENGLISH IFU:

**Caution:** This product contains natural rubber latex and other components which may cause allergic reactions, including anaphylactic responses. Isolated cases of such reactions to natural rubber latex or powder have been reported. If you experience a reaction to this product, discontinue use immediately and consult your physician. After use, wearers should visually check the gloves and remove any contamination from the outer surface before removing the gloves from the hands. Alternatively, carefully peel the gloves off the hands so that the contaminated outer surface does not touch your skin.

**Intended-use:** non-sterile examination gloves intended for use in hospitals to protect patients and users from cross-contamination, intended for use by one person during a single procedure (EN455-1).

**Storage:** Store in a cool, dry place. The opened box should be shielded from exposure to direct sun or fluorescent lighting.

### DANISH IFU:

**Vigtigt:** Dette produkt indeholder naturlatex og andre komponenter som kan fremkalde allergiske reaktioner herunder anafylaktiske reaktioner. Der er rapporteret om isolerede tilfælde af sådanne reaktioner på naturlatex og pulver. Hvis du oplever en reaktion på dette produkt, skal du straks ophøre med at bruge det og kontakte din læge.

Lav altid et visuelt tjek af handsken efter brug og fjern om nødvendigt rester fra overfladen før handskerne tages af hænderne. Alternativt bør handsken tages af, så ydersiden af handsken ikke rører ved huden på hånden. Handsken har til formål at blive benyttet af sundhedsfagligt personale til medicinske undersøgelser, for at undgå smitte mellem det sundhedsfaglige personale og patienten.

**Opbevaring:** Opbevares på et tørt og køligt sted. Når boksen er åbnet, bør den afskærmes fra direkte sollys og fluorescerende lys.