

MEDECO B.V.

Product Specifications

Brand : Klinion® Protection → SENSITIVE (LONG) SENSITIVE ULTRA SAFE

Product : Nitrile, powder-free, low dermatitis potential, non-sterile, examination gloves

Packaging & Product Reference

| Product Reference | Description | Size | Color | Box Dimension (mm)* | Piece(s) per Box | Outer Carton Dimension (mm)* | Box(es) per Outer Carton |
|-------------------|--|------|--------|---------------------|------------------|------------------------------|--------------------------|
| 102470 | Klinion Personal Protection SENSITIVE Nitrile Exam Glove | XS | Indigo | 235 * 110 * 75 | 150 | 390 * 230 * 245 | 10 |
| 102471 | Klinion Personal Protection SENSITIVE Nitrile Exam Glove | S | Indigo | 235 * 110 * 75 | 150 | 390 * 230 * 245 | 10 |
| 102472 | Klinion Personal Protection SENSITIVE Nitrile Exam Glove | M | Indigo | 235 * 110 * 75 | 150 | 390 * 230 * 245 | 10 |
| 102473 | Klinion Personal Protection SENSITIVE Nitrile Exam Glove | L | Indigo | 235 * 110 * 75 | 150 | 390 * 230 * 245 | 10 |
| 102474 | Klinion Personal Protection SENSITIVE Nitrile Exam Glove | XL | Indigo | 235 * 110 * 75 | 150 | 390 * 230 * 245 | 10 |
| 102488 | Klinion Personal Protection SENSITIVE LONG Nitrile Exam Glove | XS | Indigo | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 102484 | Klinion Personal Protection SENSITIVE LONG Nitrile Exam Glove | S | Indigo | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 102485 | Klinion Personal Protection SENSITIVE LONG Nitrile Exam Glove | M | Indigo | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 102486 | Klinion Personal Protection SENSITIVE LONG Nitrile Exam Glove | L | Indigo | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 102487 | Klinion Personal Protection SENSITIVE LONG Nitrile Exam Glove | XL | Indigo | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 102534 | Klinion Personal Protection SENSITIVE LONG Nitrile Exam Glove | XS | Blue | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 102535 | Klinion Personal Protection SENSITIVE LONG Nitrile Exam Glove | S | Blue | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 102536 | Klinion Personal Protection SENSITIVE LONG Nitrile Exam Glove | M | Blue | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 102537 | Klinion Personal Protection SENSITIVE LONG Nitrile Exam Glove | L | Blue | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 102538 | Klinion Personal Protection SENSITIVE LONG Nitrile Exam Glove | XL | Blue | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 165140 | Klinion Personal Protection SENSITIVE ULTRA SAFE LONG Nitrile Exam Glove | XS | Blue | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 165141 | Klinion Personal Protection SENSITIVE ULTRA SAFE LONG Nitrile Exam Glove | S | Blue | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 165142 | Klinion Personal Protection SENSITIVE ULTRA SAFE LONG Nitrile Exam Glove | M | Blue | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 165143 | Klinion Personal Protection SENSITIVE ULTRA SAFE LONG Nitrile Exam Glove | L | Blue | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |
| 165144 | Klinion Personal Protection SENSITIVE ULTRA SAFE LONG Nitrile Exam Glove | XL | Blue | 270 * 125 * 80 | 150 | 420 * 260 * 280 | 10 |

*Sizes are approximate

MEDECO B.V.

Product Specifications

Product description

Disposable, latex-free, powder-free examination gloves made of nitrile butadiene rubber with no Sulphur and/or accelerators. Klinion Protection SENSITIVE nitrile accelerator-free gloves have a low-dermatitis potential known to reduce potential risk for users sensitive to chemical additives.

Intended use

The Klinion Protection examination gloves are intended to be worn on the hand for medical purposes and to help prevent contamination between patient and examiner.

User Info available at www.medeco.org

Classification

MDD 93/42/EEC Class I

PPE Regulation (EU) 2016/425 Category III

→ EU declaration of conformity at www.medeco.org

Features

- Economical and hygienic protection for hands
- Ambidextrous, suitable for left and right hand
- Finger-textured surface
- Low dermatitis potential mitigating potential risk for sensitive users
- Powder-free

Product components

- Acrylonitrile Butadiene Rubber (NBR)
- Products listed on this specification are:
 - Latex-free
 - Phthalate-free
 - Absent of human or animal tissue
 - Absent of blood or blood derivative

MEDECO B.V.

Product Specifications

Specifications

Dimensions (mm)

| | | |
|-----------|---|-----------|
| Width | XS | 75 ± 5 |
| | S | 85 ± 5 |
| | M | 95 ± 5 |
| | L | 106 ± 5 |
| | XL | 116 ± 5 |
| Thickness | Cuff ref | Min. 0.05 |
| | Palm | Min. 0.06 |
| | Finger | Min. 0.08 |
| Length | Standard Cuff (Ref 102470 - 102474) | Min. 240 |
| | Long Cuff (Ref 102484-102488) Long Cuff (Ref 102534-102538) Long Cuff (Ref 165140-165144) | Min. 295 |
| | | |

Physical Properties

| | | |
|--------------------|---------------------|-------------|
| Unaged | Tensile Strength | Min. 14 MPa |
| | Elongation at Break | Min. 500 % |
| | Force at Break | Min. 6.0 N |
| Aged | Tensile Strength | Min. 14 MPa |
| | Elongation at Break | Min. 400 % |
| | Force at Break | Min. 6.0 N |
| Freedom from Holes | AQL 1.5 | |
| Powder Residue | Max 2 mg/glove | |

MEDECO B.V.

Product Specifications

Quality Standards

- Manufactured in accordance with Quality Assurance ISO 13485 and ISO 9001

In compliance with:

- EN 420:2003 + A1:2009 (fit for special purpose)
- EN ISO 374-1:2016, EN 374-2:2015, EN 374-4:2013, EN ISO 374-5:2016
- EN 455-1:2000, EN 455-2:2015, EN455-3:2015, EN455-4:2009
- EN 16523-1:2015

Sterilization

N/A - Product is provided non-sterile

Storage conditions

Store in a cool and dry place. Opened boxes should be kept away from fluorescent and sunlight. Gloves are packed in dispenser which is suitable for transport. Keep the gloves in the box when not in use.

Shelf life

3 years

Pictures



Ref. product 102472





Ref. product 102487



Ref. product 102535



Ref. product 165143

| Approval | Name | Date | Signature |
|-----------------|------------------|------------|---|
| Product Manager | Levine van Buren | 20-12-2018 |  |
| RA-Manager | Francis Dekker | 20-12-2018 |  |