

EC Declaration of Product Conformity

Unicare Soft Nitrile (GS003* & GS004*)

Non-sterile, powder free, nitrile single use glove

The manufacturer established in the Community declares that the new medical device and PPE described hereafter, Unicare Soft Nitrile, is in conformity with the provisions of Council Directive 93/42/EEC as amended, national standards EN455-1:2000, EN455-2:2015, EN455-3:2015 and EN455-4:2009 and is Class I self-certified and

is in conformity with the provisions of Council Directive 89/686/EEC as amended and where such is the case, with national standards transposing harmonized standards EN374-1:2003, EN374-2:2003, EN374-3:2003 and EN420:2003+A1:2009

is identical to the PPE which is the subject of EC certificate of conformity no. 8590 issued by SATRA Technology Centre, Wyndham Way, Telford Way, Kettering, Northamptonshire, NN16 8SD, UK (Notified Body No. 0321)

conforms to the provisions of Council Directive 89/686/EEC Article 11A under the supervision of the Notified Body, SATRA Technology Centre, (Notified Body No. 0321)

Unigloves are a market-leading manufacturer of premium quality, specialist, single use gloves.

We make and supply over 2 billion gloves annually to medical professionals, as well as specialists in the food, automotive, janitorial and care sectors.

 **Unigloves**
Lakeside Park
Neptune Close
Rochester
Kent
ME2 4LT
 01634 726 516
 unigloves.co.uk

Product Information

Glove Care: Store between 5°C and 25°C in a cool dry environment away from direct sunlight.

Food Contact: Unicare Soft Nitrile gloves have been tested in accordance with EN1186. They are suitable for use with all food categories, in situations of short term repeat contact.

User Information

Cleaning: Gloves cannot be cleaned and reused. Before use, check gloves for physical damage. Do not use damaged gloves.

Fitting and sizing: Only wear products of a suitable size. Products which are either too loose or tight will restrict movement and will not provide the optimum level of protection.

Allergy advice: This product does not contain natural rubber latex but may contain low levels of residual chemical accelerators, which may cause allergic reactions.



Available sizes (100s)

| | |
|-------------|--------|
| Extra Small | GS0031 |
| Small | GS0032 |
| Medium | GS0033 |
| Large | GS0034 |
| Extra Large | GS0035 |

Available sizes (200s)

| | |
|-------------|--------|
| Extra Small | GS0041 |
| Small | GS0042 |
| Medium | GS0043 |
| Large | GS0044 |
| Extra Large | GS0045 |



Chris Wahlers

Director

Unigloves (UK) Limited