

Certificate of Analysis / Shipment Inspection

Brand : SRITRANG 14000060
Product : STGTH-10X100-NOFSOG-THD-BLU-S-NS-BX
Glove Spec : NOOOGF-S-BLU-TH-BD-NS
Size : S
Delivery quantity : 200 CTN
Delivery Date : 08.05.2024
Delivery Order No. : 4940013893
Customer Order No. : 2024/05139
Batch / Lot No. : 3893
MFG. Date : 042024
EXP. Date : 042027

Sampling plan acc. to ISO 2859-1	Specification	Test Result	Sampling Procedure
Barrier properties			
No. of gloves tested	AQL ≤ 2.5	200	INSPECTION LEVEL : GI ,SINGLE NORMAL
No. of Defect		0	INSPECTION LEVEL : GI ,SINGLE NORMAL
Accept/ Reject		AC	
Major defects			
No. of gloves tested	AQL ≤ 2.5	200	INSPECTION LEVEL : GI ,SINGLE NORMAL
No. of Defect		0	INSPECTION LEVEL : GI ,SINGLE NORMAL
Accept/ Reject		AC	
Minor defects			
No. of gloves tested	AQL ≤ 4.0	200	INSPECTION LEVEL : GI ,SINGLE NORMAL
No. of Defect		0	INSPECTION LEVEL : GI ,SINGLE NORMAL
Accept/ Reject		AC	
Dimension			
Width (mm)	80+/-10	84	ACCORDING TO TIS 1056
Length (mm)	MIN 220	246	ACCORDING TO TIS 1056
Thickness Finger in mm (SW)	>/=0.11	0.12	ACCORDING TO TIS 1056
Thickness Palm in mm (SW)	>/=0.08	0.09	
Tensile Properties			
Elongation in %, unaged	>/=500	580	ACCORDING TO EN 455-2
Elongation in %, aged	>/=400	521	
Biocompatible			
Powder Content (mg/glove)	</=2.0	0.1	ACCORDING TO ASTM D6319

Approved by QA: Proifon Rujiraworrarak

Proilfon Rujiraworrarak
Date 10.05.2024

DOCUMENT CENTER

Brand

Product

Glove Spec

Size

Delivery quantity

Delivery Date

Delivery Order No.

Customer Order No.

Batch / Lot No.

MFG. Date

EXP. Date

: SRITRANG 14000061

: STGTH-10X100-NOFSOG-THD-BLU-M-NS-BX

: NOOOGF-S-BLU-TH-BD-NS

: M

: 600 CTN

: 08.05.2024

: 4940013893

: 2024/05139

: 3893

: 042024

: 042027

Sampling plan acc. to ISO 2859-1	Specification	Test Result	Sampling Procedure
Barrier properties			
No. of gloves tested	AQL ≤ 2.5	315	INSPECTION LEVEL : GI ,SINGLE NORMAL
No. of Defect		4	INSPECTION LEVEL : GI ,SINGLE NORMAL
Accept/ Reject		AC	
Major defects			
No. of gloves tested	AQL ≤ 2.5	315	INSPECTION LEVEL : GI ,SINGLE NORMAL
No. of Defect		0	INSPECTION LEVEL : GI ,SINGLE NORMAL
Accept/ Reject		AC	
Minor defects			
No. of gloves tested	AQL ≤ 4.0	200	INSPECTION LEVEL : GI ,SINGLE NORMAL
No. of Defect		0	INSPECTION LEVEL : GI ,SINGLE NORMAL
Accept/ Reject		AC	
Dimension			
Width (mm)	95+/-10	95	ACCORDING TO TIS 1056
Length (mm)	MIN 230	244	ACCORDING TO TIS 1056
Thickness Finger in mm (SW)	>/=0.11	0.13	ACCORDING TO TIS 1056
Thickness Palm in mm (SW)	>/=0.08	0.09	
Tensile Properties			
Elongation in %, unaged	>/=500	581	ACCORDING TO EN 455-2
Elongation in %, aged	>/=400	527	
Biocompatible			
Powder Content (mg/glove)	</=2.0	0.1	ACCORDING TO ASTM D6319

Brand

: SRITRANG 14000062

Product

: STGTH-10X100-NOFSOG-THD-BLU-L-NS-BX

Glove Spec

: NOOOGF-S-BLU-TH-BD-NS

Size

: L

Delivery quantity

: 500 CTN

Delivery Date

: 08.05.2024

Delivery Order No.

: 4940013893

Customer Order No.

: 2024/05139

Batch / Lot No.

: 3893

MFG. Date

: 042024

EXP. Date

: 042027

Sampling plan acc. to ISO 2859-1	Specification	Test Result	Sampling Procedure
Barrier properties			
No. of gloves tested	AQL ≤ 2.5	200	INSPECTION LEVEL : GI ,SINGLE NORMAL
No. of Defect		0	INSPECTION LEVEL : GI ,SINGLE NORMAL
Accept/ Reject		AC	
Major defects			
No. of gloves tested	AQL ≤ 2.5	200	INSPECTION LEVEL : GI ,SINGLE NORMAL
No. of Defect		0	INSPECTION LEVEL : GI ,SINGLE NORMAL
Accept/ Reject		AC	
Minor defects			
No. of gloves tested	AQL ≤ 4.0	200	INSPECTION LEVEL : GI ,SINGLE NORMAL
No. of Defect		0	INSPECTION LEVEL : GI ,SINGLE NORMAL
Accept/ Reject		AC	
Dimension			
Width (mm)	110+/-10	113	ACCORDING TO TIS 1056
Length (mm)	MIN 230	243	ACCORDING TO TIS 1056
Thickness Finger in mm (SW)	>/=0.11	0.11	ACCORDING TO TIS 1056
Thickness Palm in mm (SW)	>/=0.08	0.09	
Tensile Properties			
Elongation in %, unaged	>/=500	597	ACCORDING TO EN 455-2
Elongation in %, aged	>/=400	542	
Biocompatible			
Powder Content (mg/glove)	</=2.0	0.1	ACCORDING TO ASTM D6319