DIGITAL THERMOMETER

MD

OPERATING INSTRUCTIONS

REF: SC0006

PLEASE READ CAREFULLY BEFORE USING

The Digital Thermometer provides quick, highly accurate reading of your temperature. Please read the instructions thoroughly. They are added to acquaint you with features of your new thermometer so that you can enjoy its functions and years of dependable performance. If the thermometer is used by a child, the supervision of the adult is compulsory. This thermometer is manufactured according to the system of EN13729-3:2000, ASTM and complies with the requirements of IEC 60601-1:1988+A1+A2, IEC 60601-1-2:2001. And the manufacturer is ISO 9001:2000 and ISO 13485:2003 certified.

PRODUCT DESCRIPTION



SPECIFICATIONS

Range:	32.0℃~42.9℃ (90.0℉~109.9℉) (℃/℉ chosen by manufacturer)
Accuracy:	±0.1℃ (±0.2℉) (32.0℃~34.9℃) (89.6℉~107.6℉)
	±0.2℃ under 35.5℃ or over 42.0℃
	±0.3℃ (±0.6℉) under 95.9℉ or over107.6℉)
	at standard room temperature of 25℃ (77.0℉)
Display:	Liquid crystal display, 3 1/2 digits
Memory:	For storing the last measured value
Power:	Two 1.5V button size battery (model SR41, UCC 392 or LR41)
Power consumption:	0.15 mW/min in measurement mode
Battery life:	Approx 200 hours
Dimension:	12.4cm x 1.8cm x 0.9cm (L x W x H)
Weight:	Approx. 11 grams including battery
Using environment:	Temperature: 10℃~40℃ (50℉~104℉)
	Relative humidity: 25%RH~80%RH
Storage and transportation condition:	Temperature: -25℃~55℃ (-13℉~131℉)
	Relative humidity: 25%RH~80%RH

HOW TO USE

1. Wipe or dip the probe with ethyl alcohol for disinfection.
2. Push the On/Off button. The LCD will show 88.8 for 2 seconds and you will hear a beep simultaneously. Last tested temperature will be shown for about 2 second. Then LCD shows the current temperature reading. If the last tested temperature is higher than Lo℃ (Lo℉) sign shows.
3. Position the probe in mouth, rectum, armpit etc.
4. Once the temperature-taking is finished, the degree sign of ℃ (℉) on the LCD will stop flashing, and you will hear beeps for approx. 10 seconds. The reading will not change. The tested temperature is shown on LCD.
During or after the test, if the test temperature is < 32.0℃ (or 90.0℉), the LCD shows Lo℃ (or Lo℉). If the tested temperature is extra high, the LCD shows Hi℃ (or Hi ℉). Consult doctor at once.
During or after the test, if the tested temperature is >37.8℃(100.0℉), the buzzer sounds "Di-Di-Di-Di-", if the tested temperature is ≤37.8℃(100.0℉), the buzzer whose frequency are much higher are "Di-Di-Di——Di-Di-Di——Di-Di-Di".
5. The thermometer will automatically turn off in about 10 minutes for energy-saving. However, to prolong battery life, it is suggested to press the On/Off button to turn the thermometer off once the test is finished.
6. Please dispose exhausted battery and thermometer according to the relevant law.

Note: In step 2, if the room temperature is higher than 32.℃(90℉), LCD will display the room temperature, instead of Lo℃ (or Lo℉).
Note: Use of the probe cover may result in a 0.1℃(0.2℉) discrepancy from actual temperature.

ORAL USE

Put the probe into either position marked by " ✓ " as showed in the picture right-hand for about 1 minute. To ensure the accuracy, please keep the mouth close for 5 minutes before taking temperature. The normal temperature by this methods is 35.7℃-37.5℃(96.3℉-99.5℉).

RECTAL USE

Lubricate the probe with the petroleum jelly for easy insertion. Gently insert the probe about one centimeter then that 1.5 inch into the patient's bottom and hang for about 1 minute. The normal temperature by this methods is 36.6℃-37.8℃(98℉-100℉).

AXILLARY USE

From a medical viewpoint, this method will always provide inaccurate readings. Wipe armpit with a dry towel. Place probe in the armpit and hold it firmly against the body side for about 1.5 minute. The normal temperature by this methods is 35.5℃-37.1℃(96-98.8℉).

CLEANING AND DISINFECTION

1 Clean thermometer by wiping it with a dry cloth and sterilize the probe with ethyl alcohol.
2 Do not immerse thermometer into any chemical thinner.
3 Not completely waterproof. Do not soak the LCD in water. Only the probe is waterproof.

PRECAUTION

* Do not drop thermometer. It is not shock proof.
* Do not bend thermometer.
* Do not boil the probe.
* Do not store thermometer in direct sunlight.
* Do not attempt to disassemble thermometer except replacing the battery.

BATTERY REPLACEMENT

1. When the ▮ appears or the lower right corner of the LCD, the battery is exhausted and needs replacement.
2. Pull the battery cover off with hand in the direction shown below.
3. Remove the old battery and replace with new one (model number: SR41). Polarity: 1 on.
4. Use pointed object such as pen to take out the exhausted battery. Discard exhausted battery out each of children.
5. Place a new 1.5V D.C. button size battery type SR41, UCC 392 or LR41 or equivalent in the chamber with positive side faced up and negative side faced down.
6. Cover the battery.

LIMITED WARRANTY

The thermometer is guaranteed for 90 days from the date of purchase. If the thermometer does not function properly due to defective components or poor workmanship, we will repair or replace it free of charge. All components are covered by this warranty excluding the battery. The warranty does not cover damages to your thermometer due to improper handling.

EC REP

ISTRUZIONI D' USO ALGIORNO E IL / OPERATING INSTRUCTION FIRLIED ON 2010-07

TERMOMETRO DIGITALE

ISTRUZIONI

SI PREGA DI LEGGERE ATTENTAMENTE PRIMA DELL'UTILIZZO

Il termometro digitale fornisce una valiore ed accurata lettura della vostra temperatura. Si prega di leggere attagchiatente le istruzioni. Se il termometro è usato ed utilizzato dcon bambino. Se aspetti come di un adulto è ancessaria. Questo termometro è stato costruito nel rispetto del sistema di qualità EN13729-3:2000, ASTM ed è conforme ai requisiti IEC 60601-1:1988+A1+A2, IEC 60601-1-2:2001. Inoltre il fabbricante è certificato ISO 9001:2000 e ISO 13485:2003.

DESCRIZIONE DEL PRODOTTO



SPECIFICHE TECNICHE

Range:	32.0℃~42.9℃ (90.0℉~109.9℉) (℃/℉ chosen by manufacturer)
Accuracy:	±0.1℃ (±0.2℉) (32.0℃~34.9℃) (89.6℉~107.6℉)
	±0.2℃ under 35.5℃ or over 42.0℃
	±0.3℃ (±0.6℉) under 95.9℉ or over107.6℉)
	at standard room temperature of 25℃ (77.0℉)
Display:	Liquid crystal display, 3 1/2 digits
Memory:	For storing the last measured value
Power:	Two 1.5V button size battery (model SR41, UCC 392 or LR41)
Power consumption:	0.15 mW/min in measurement mode
Battery life:	Approx 200 hours
Dimension:	12.4cm x 1.8cm x 0.9cm (L x W x H)
Weight:	Approx. 11 grams including battery
Using environment:	Temperature: 10℃~40℃ (50℉~104℉)
	Relative humidity: 25%RH~80%RH
Storage and transportation condition:	Temperature: -25℃~55℃ (-13℉~131℉)
	Relative humidity: 25%RH~80%RH

MODALITA' D'UTILIZZO

1 Strofinare o immergere il sensore nell'alcool per una rapida disinfezione
2 Premere il tasto ON/OFF. Sull'LCD apparirà 88.8° per 2 secondi seguito da un segnale sonoro (bep). L'ultima temperatura rilevata verrà visualizzata per circa 2 secondi. Successivamente comparirà sull'LCD la temperatura "ture" in 37.0°.
3 Posizionare il termometro in bocca, retto, ascella
4 Una volta che la temperatura sarà misurata, i valori visualizzati sull'LCD smetteranno di lampeggiare e si potrà sentire un segnale sonoro per circa 10 secondi. La temperatura potrà quindi essere visualizzata sull'LCD. Durante o dopo la misurazione della temperatura, se la temperatura (< 32.0℃(o 90.0℉) sull'LCD comparirà Lo℃(o Lo℉); se la temperatura è troppo alta, sull'LCD comparirà lo scritto Hi℃ (o Hi ℉). Quindi consultare un medico. Dopo aver misurato la temperatura se essa ≥ 37.8℃(100.0℉), il segnale sonoro sarà "Di-Di-Di-Di-"; se la temperatura sarà ≤37.8℃(100.0℉), la frequenza del segnale sonoro sarà più alta "Di-Di-Di——Di-Di-Di——Di-Di-Di"
5 Il termometro si spegnerà automaticamente dopo circa 10 minuti, per prolungare la durata della batteria, si consiglia di premere il pulsante ON/OFF una volta che si smette l'utilizzo del termometro.
6 Si prega di smaltire le batterie esaurite e lo stesso termometro secondo la leggi vigenti.

NOTA: Nel passo 2, se la temperatura della stanza è più alta di 32.0℃ (90℉), sull'LCD comparirà la temperatura della stanza, invece di Lo℃ (o Lo℉).
Nota: Il coprendi il sensore dà sulla accuratezza della temperatura, si potrebbe avere una discrepanza de 0.1℃ (0.2℉) riguardo alla temperatura reale.

USO ORALE

Inserire la sonda nella posizione indicata dal simbolo " ✓ " come indicato nella figura a destra e tenerla in posizione. Per ottenere carte, tenere la bocca chiusa per 5 minuti prima di procedere la temperatura. La temperatura ,con questo metodo- è di 35.7℃-37.5℃(96.3℉-99.5℉).

USO RETTALE

Lubrificare il sensore con della vaselina per una miglior inserimento. Inserire il sensore per circa 1cm (profondità 1.5 inch) nel paziente e tenerlo in posizione per circa 1 minuto. La temperatura normale misurata con questo metodo è di 36.6℃-37.8℃(98℉-100℉).

USO ASCELLARE

Da un punto di vista medico, questo metodo porterà sempre ad una lettura inaccurata. Asciugare l'ascella con un panno asciutto. Posizionare la sonda nell'ascella e premere contro il corpo del paziente per circa 1.5 minuti. La temperatura normale con questo metodo è 35.5℃-37.1℃(96-98.8℉).

PULITURA E DISINFEZIONE

1 pulire il termometro ai vapori asciutto e sterilizzarlo con alcool etilico
2 Non immergere il termometro a contatto con agenti chimici
3 Non completamente impermeabile. Non immergere il display in acqua. Solo il sensore è impermeabile

PRECAUZIONI

* non far cadere il termometro
* non piegare il termometro
* non bollire il sensore
* non esporre il termometro alla luce diretta del sole
* non tentare di smontare il termometro, eccetto per la sostituzione della batteria

SOSTITUZIONE BATTERIA

1 Quando il ▮ appare nell'angolo in basso a destra dell'LCD, la batteria è scarica ed il termometro sostituito
2 Tirare il coperchio copri batteria sulla direzione sotto indicata
3 Tenere delicatamente la base dove è alloggiata la batteria di cattiva lega
4 Usare un oggetto appuntito come una penna per togliere la batteria scarica
5 Inserire la nuova pila da 1.5V, con il lato + rivolto verso l'alto e il lato – rivolto verso il basso
6 Rimettere il coperchio copri batteria

GARANZIA

Il termometro è garantito per il periodo previsto dalla legge vigente. La garanzia è valida solo se l'apparecchio non è stato manomesso o se è stato utilizzato in modo corretto secondo le istruzioni d'uso. Tutti i componenti sono coperti dalla garanzia ad esclusione della batteria

EC REP

正面

反面