

Date: September 24, 2021

Dear Valued customer,

Certificate of Analysis

Product Specification Code : LXXOGF-S-WHT-TH-BD-NS
Product Name : Latex Examination Gloves, Powdered, Non Sterile
Surface : Textured at Finger Tip
Glove Color : WHITE

Characteristics	Specification LXXOGF-S-WHT-TH-BD-NS	Average Actual Result
1.Length (mm.)		
- XS	Min 220	246
- S	Min 220	246
- M	Min 230	245
- L	Min 230	246
- XL	Min 230	No Produce
2. Width (mm.)		
- XS	≤ 80	79
- S	80+/-10	84
- M	95+/-10	95
- L	110+/-10	108
- XL	≥ 110	No Produce
3.Single Wall Thickness (mm.)		
- Finger	Min 0.11	0.13
- Palm	Min 0.08	0.11
4. Physical Properties		
4.1 Before Aging		
- Force (N)	Min 7.0	10.2
- Elongation (%)	Min 650	808
4.2 After Aging		
- Force (N)	Min 6.0	9.7
- Elongation (%)	Min 500	773
5. Hole	AQL ≤ 2.5	

Remark: 1. Average test results in January – August 2021

Best Regards,



Ms. Sureerat Choosri

Product Manager (Glove)