

# SRI TRANG GLOVES (THAILAND) PUBLIC COMPANY LIMITED

10 Soi 10 Phetkasem Road, Hat Yai, Hat Yai, Songkhla 90110 Thailand  
Tel. (66) 074-344-663 Fax. (66) 074-344-677

SCT.PD.SP.05.219-140222 R.03

DAR-SCT-2022-0213

PRODUCT SPECIFICATION SHEET : NOPOGT-H-EU-M-NS

PAGE 1 of 1

## GLOVE SPECIFICATION

<b>Product Specification Group Code :</b>	<b>NOPOGT-H-EU-M-NS</b>
<b>Product Code :</b>	<b>DNOTHOG (With Color)</b>
<b>Product name :</b>	<b>Nitrile Examination Gloves, Powder Free, Chlorinated, Non Sterile</b>
<b>Grade :</b>	<b>EU Spec, Medical Grade</b>

### GENERAL INFORMATION

<b>Type :</b>	Single-use disposable examination or protective glove
<b>Marking :</b>	Gloves are not marked to designated size.
<b>Shape :</b>	Ambidextrous
<b>Material :</b>	Nitrile Butadiene Rubber
<b>Cuff :</b>	Rolled cuff
<b>Surface :</b>	Textured

<b>Thickness :</b> (single wall according to ASTM D6319, ASTM D3767)	<b>Finger :</b> Min. 0.05mm	<b>Palm :</b> Min. 0.05mm
---	-----------------------------	---------------------------

<b>Residual powder :</b> (according to EN 455-3, EN ISO 21171)	Max. 2.0mg/glove	<b>Force at break :</b> (according to EN 455-2)	Before Aging : Med. $\geq$ 6.0N After Aging : Med. $\geq$ 6.0N
---	------------------	--	---

<b>Length :</b> (according to EN 455-2)	XS-XL : Med. $\geq$ 240mm	<b>Tensile Strength :</b> (according to ASTM D6319, ASTM D412, ASTM D573)	Before Aging : Min. 14Mpa After Aging : Min. 14Mpa
--	---------------------------	--	---


<b>% Elongation :</b> (according to ASTM D6319, ASTM D412, ASTM D573)	Before Aging : Min. 500% After Aging : Min. 400%
--	---

<b>Available Sizes &amp; Palm width (mm) :</b> XS (5-6) : $\leq$ 80, S (6-7) : 80+/-10, M (7-8) : 95 +/-10, L (8-9) : 110+/-10, XL (9-10) : $\geq$ 110 (according to EN 455-2)
---

**Storage:** 3 years storage life (date of expiration see on the box). Keep area cool, dry and dust free, avoid ventilation and storage close to photocopy equipment. Copper-ions discolor the glove. Protect gloves against ultraviolet light sources, as sunlight and oxidizing agents. Storage above 30°C will lead to accelerated aging and should be avoided.

### STANDARDS, GUIDELINES AND QUALITY CERTIFICATION

(Every mentioned standard is used in the latest edition.)

<b>Certification :</b>	ISO 9001, ISO 13485		
<b>Conformity to guidelines :</b>	Medical Device Regulation (EU)2017/745 : Class I under Rule1 and Rule 5 according to Annex VIII (Subject to EU Declaration of Conformity to brand owner) Personal Protective Equipment Regulation (EU) 2016/425 : Category I or Category III (Subject to EU Type-Examination Certificate)		
<b>Conformity to standards :</b>	<u>Medical</u> EN 455 ASTM D6319	<u>PPE</u> CAT I EN ISO 21420	<u>PPE</u> CAT III EN ISO 21420 EN 374 EN 16523-1
<b>Conformity to food regulation :</b>	1935/2004		

### WARNING AND REPORTING SYSTEM

<b>Potentially allergenic ingredients :</b>	Dithiocarbamate type. (For further information, the actual version of the ingredients list is given on request.)
---	---

Please report any incidents with the products to [STGT\\_TA@sritranggroup.com](mailto:STGT_TA@sritranggroup.com) and [regulatory@sritranggroup.com](mailto:regulatory@sritranggroup.com) without undue delay. Our Product Safety Officers will review all information provided and might contact the original informant for further clarification.